

Mona Lisa N.V.

Kapelstraat 1, 3540 Herk-de-Stad, Belgium

Device Identification:

Sterile Mona Lisa® Cu375, Multi-Safe® CU375, Mithra®-Load 375, Mona Lisa® Cu375 SL, Multi-Safe® CU375 Short, Mona Lisa® ST Cu300, Mithra®-Flex 300 and CU-Safe® T300 Intrauterine devices.

All the above IUD's are supplied with or without Hysterometer size 12 in a separate sterile pouch

Intended Purpose of Device:

Copper-containing intrauterine devices intended for long-lasting reversible female contraception

has been assessed and certified as meeting the requirements of

Directive 93/42/EEC

on Medical Devices, Annex II, Section 4

It is certified that the manufacturer's design dossier (and product, where applicable) for the above device has been examined and, based on the evidence submitted, it is considered that the device conforms to the relevant Essential Requirements of EC Directive 93/42/EEC.

This certificate is issued in conjunction with a certificate covering the full quality assurance system to Annex II, which must be subject to satisfactory surveillance audits.

This certificate is valid from 22 May 2020 until 07 August 2022
Issue 2. Certified since 08 August 1997
and first certified by SGS Belgium NV since 16 December 2019

Certification is based on reports numbered BE/AND 201993


Authorised by

SGS Belgium NV, Notified Body 1639

SGS House Noorderlaan 87 2030 Antwerp Belgium
t +32 (0)3 545-48-48 f +32 (0)3 545-48-49 www.sgs.com

LPMD5007 - Certificate CE1639 Annex II-4_EN rev. 02

Page 1 of 2



This document is issued by the Company subject to its General Conditions of Certification Services, unless otherwise agreed, accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein. The authenticity of this document may be verified at <https://www.sgs.com/en/certified-clients-and-products/certified-client-directory>. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Mona Lisa N.V.

Directive 93/42/EEC on Medical Devices, Annex II, Section 4

Issue 2

Certification is based on reports numbered BE/AND 201993 dated 29th November 2017

Addenda to that report have been issued on the following dates:

<u>Addendum Date</u>	<u>Reason for Addendum</u>
25 May 2018	Device description and intended purpose amended for consistency
13 November 2018	The use of DuPont Tyvek® (MPTP) transition material
12 August 2019	Change of the address
08 January 2020	Change to labelling to included specific reference to the product including a Size 12 Hysterometer
06 May 2020	Change of supplier for the polyamide (Nylon) withdrawal thread Validation of a dedicated load ethylene oxide sterilization cycle at Sterigenics (Petit Rechain)

